

Republic of the Philippines
Province of Davao del Norte
 Government Center, Mankilam, Tagum City

PR Number : 1306-1545

Bid Number : B20130060

Date : July 30, 2013

Page : 1

Item No.	Qty./Unit	Item	Bid Price	
			Unit Price	Total Amount

1 1 unit Cautery Machine

Specifications:

Standards Accessories:

- Twin foot pedal for Monopolar & bipolar use
- Second pedal for independent bipolar use only
- Cable for bipolar forceps with Euro-forceps
- Adapter for monopolar outlet for plug size 2-8mm
- Bipolar Tissue forceps with 18cm, tip 1.0mm
- Monopolar cable for Laparoscopy
- Adapter for USA-type neutral electrode
- Adapter for Valleylab.Conmed-type neutral electrode
- Cable for connecting neutral electrodes

Pencil

Return electrode adult

AVR (2.0KVA)

Cart

2 1 unit Patient Monitor (Multi Parameter Patient Monitor)

Specifications:

-Brand new tech

- Function of Alarm event reviewing can store 1800 peices of alarm events
- Function of NIBP measurement reviewing, can store 1000 pieces of NIBP
- Function of reviewing 30 minutes one important lead's ECG waveform
- long life time backlight buttons
- Parameter display with big character
- Function of display of short trend
- Portable design, stylish and convenient

Purchase of Medical Equipments for the use of DDNH-Carmen

APPROVED BUDGET FOR THE CONTRACT (ABC) : P 500,000.00

PGSO Warehouse

Period of Delivery : _____ day/s

**SUPPLIERS MUST SPECIFY/INDICATE
 BRAND NAMES UPON QUOTATION**

INSTRUCTIONS IN FILLING UP THIS BID FORM

Bids Distributed To:

1. Bidder must read carefully the statement at the back hereof.
2. Bid Quotation must be typewritten or printed in blue or black ink.
3. Unit price, total amount and grand/lot price must be properly indicated.
4. Delivery period must be indicated in the Bid form unless otherwise provided by the procuring unit/office. Failure to indicate the period of delivery shall automatically mean 10 days delivery after receipt of the Purchase Order.
5. Place of Delivery shall be at the PGSO Warehouse unless otherwise indicated in the bid form.
6. Submit Bid in the sealed envelope.
7. Unprinted name of establishment, unprinted name of the bidder and unsigned bid shall be disqualified.
8. Erasures/alterations should be initialed accordingly.

.....
 Name of Establishment

.....
 Name of Bidder

.....
 Signature of Bidder

Republic of the Philippines
Province of Davao del Norte
Government Center, Mankilam, Tagum City

PR Number : 1306-1545

Bid Number : B20130060

Date : July 30, 2013

Page : 2

Item No.	Qty./Unit	Item	Bid Price	
			Unit Price	Total Amount

- 15" TFT LCD display, high resolution (1024*768)
- Rechargeable lithium Battery
- Unique all-lead ECG on-one-screen display, which can facilitate the diagnosis
- Can store the trend data for 168 hours and has the function of displaying trend data
- Rechargeable maintenance-free battery, can continue working when AC power is off
- Can be connected with the central unit to realize centralizing monitoring
- easy to operate at night dual alarm light
- Nurse call function
- Complete built-in module design ensures stable and reliable performance
- Central Monitoring ready
- Is resistant to high-frequency electrode and is protected against defibrillation effects

Remarks : All bidders must attach brochure and follow specification and attach warranty for 1 year or more.

Note:

Chargeable to PHIC-PFPR-Capitation Fund.....

* * * GRAND/LOT PRICE : P _____

Purchase of Medical Equipments for the use of DDNH-Carmen

APPROVED BUDGET FOR THE CONTRACT (ABC) : P 500,000.00

PGSO Warehouse

Period of Delivery : _____ day/s

**SUPPLIERS MUST SPECIFY/INDICATE
BRAND NAMES UPON QUOTATION**

INSTRUCTIONS IN FILLING UP THIS BID FORM

Bids Distributed To:

1. Bidder must read carefully the statement at the back hereof.
2. Bid Quotation must be typewritten or printed in blue or black ink.
3. Unit price, total amount and grand/lot price must be properly indicated.
4. Delivery period must be indicated in the Bid form unless otherwise provided by the procuring unit/office. Failure to indicate the period of delivery shall automatically mean 10 days delivery after receipt of the Purchase Order.
5. Place of Delivery shall be at the PGSO Warehouse unless otherwise indicated in the bid form.
6. Submit Bid in the sealed envelope.
7. Unprinted name of establishment, unprinted name of the bidder and unsigned bid shall be disqualified.
8. Erasures/alterations should be initialed accordingly.

.....
Name of Establishment

.....
Name of Bidder

.....
Signature of Bidder